

Institute of radiation physics

Préparation de solutions radioactives étalons par un labora- toire primaire

Claude Bailat, Thierry Buchillier

*Frédéric Juget, Teresa Duran, Youcef
Nedjadi, Corinne Moratal, Sandrine
Zufferey*

Let's step back from the end certificate to the primary measurement....



Institut de
radiophysique
appliquée
Rue du Grand-Père 1
CH-1007 Lausanne

Page

Fiche descriptive

Echantillons 1407-1, 1407-2, 1407-3 et 1407-4

Délivré à :

LABOR SPIEZ
Austraase
CH-3700 SPIEF7

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

Eckert & Ziegler
Isotope Products

CERTIFICATE OF CALIBRATION GAMMA STANDARD SOLUTION

Radionuclide: Pb-210
Half-life: 22.3 ± 0.2 years
Catalog No.: MSC-S1D
Source No.: 1989-91

Customer: ECKERT & ZIEGLER ISOTOPE PRODUCTS GMBH
P.O. No.: 37388
Reference Date: 15-Jun-09 12:00 PST
Contained Radioactivity: 9.458 μ Ci 349.9 kBq

Physical Description:

- A. Mass of solution: 1.08654 g in 1 mL V-Vol
B. Chemical form: $Pb(NO_3)_2$ in 1M HNO_3
C. Carrier content: 10 μ g Pb/mL of solution
D. Density: 1.0318 g/mL @ 20°C

Radiopurities:

None detected

Radionuclide Concentration: 8.705 μ Ci/g, 322.1 kBq/g

Method of Calibration:

This source was prepared from a weighed aliquot of solution whose activity in μ Ci/g was determined using gamma ray spectrometry.

Peak energy used for integration: 46.5 keV
Branching ratio used: 0.0418 gammas per decay

Uncertainty of Measurement:

- A. Type A (statistical) uncertainty: ± 0.3 %
B. Type B (systematic) uncertainty: ± 7.0 %
C. Uncertainty in aliquot weighing: ± 0.0 %
D. Total uncertainty at the 95% confidence level: ± 7.0 %

Notes:

- See reverse side for leak test(s) performed on this source.
- EZIP participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of radionuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (as in NRC Regulatory Guide 4.15).
- Nuclear data was taken from NCRP Report No. 58, 1985.
- This solution has a working life of 5 years.

Daniel H. DeLeon
Quality Control

22-Jul-09
Date

EZIP Ref. No.: 1989-91

ISO 9001 CERTIFIED

Medical Imaging Laboratory

24937 Avenue Tibbitts, Valencia, Ca 91355

Industrial Gaging Laboratory

1800 North Keystone Street, Burbank, California 91514



Institut universitaire de radiophysique appliquée

Page 1 sur 2

Mandaté par l'Office fédéral de métrologie METAS pour la
détermination et la diffusion de l'unité légale d'activité

Certificat d'étalonnage

No 0808-12.0311

Délivré à

AAA Advanced Accelerator Applications
Technoparc
20, Rue Diesel
01630 - Saint Genis Pouilly
France

Page 2 sur 2

Objet

1 source radioactive de F-18

Forme chimique

Fluoro-Deoxy-Glucose FDG

Identification

BP-FG-080826-B-01

Impuretés radioactives

Aucun radionucléide émetteur gamma de
longue période n'a été mesuré dans la
source. (LD < 0.01%)

its étalonnée en 1997

Activité Par unité de
masse

(607.28 \pm 4.69¹) MBq/g

national de cette
des Poids et Mesures
m.org

Date de référence

26 août 2008 à 07 h 15 TU²

Le responsable de la mesure

Institut de Radiophysique Appliquée

Dr. Yves Collari

Dr. Claude Bailat
Chef du groupe de Radiométrie

Lausanne le 1er septembre 2008

¹ Le nombre qui suit le symbole \pm est l'incertitude élargie. (voir page 2)
² Correspond à 9:15 Heure CET. (Heure locale Genève, horaire d'été)

Institut universitaire de
Radiophysique Appliquée
Grand-Père 1
CH-1007 Lausanne

La publication ou la reproduction de ce certificat n'est autorisée
que dans sa forme intégrale

i, ISBN 92-67-20188-3.

Institut universitaire de
Radiophysique Appliquée
Grand-Père 1
CH-1007 Lausanne

La publication ou la reproduction de ce certificat n'est autorisée
que dans sa forme intégrale



Czech Metrology Institute
Okružní 31, 638 00 Brno

Phone: +420 545 555 111, Fax: +420 545 222 728, www.cmi.cz

Workplace: Regional Branch Prague, Radiová 1136/3, 102 00 Praha 10
Ionizing Radiation Building, Radiová 1288/1a
Phone: +420 266 020 497, Fax: +420 266 020 466

CERTIFICATE

Certificate No.: 1035 - SE - 440293-15 Type: CBSS 2 Serial No: 030215-1503033

Radionuclide	Half life, days	Activity, kBq	Combined standard uncertainty, %
Au-241	157800	1,033	1,1
Cd-109	462,6	4,383	1,5
Ce-139	137,5	0,2436	1,2
Co-57	271,26	0,2202	1,1
Co-60	1925,4	0,6277	1,1
Cs-137	11019	0,5707	1,3
Sr-113	115,1	1,742	2,2
Sr-85	64,78	2,195	1,2
Y-88	106,6	2,188	1,5
Pb-210	8108	10,66	1,5

Mass: 539.0 g Density: 0.98 ± 0.01 g/cm³ Volume: 550.0 \pm 5.5 cm³

Radionuclide impurities: gamma < 0,1 %

Reference date: 1.5.2015

Homogeneity: better than: 1 %

Description:

Radioactive material is homogeneously dispersed in silicone resin. Composition of the matrix: C - 0.324 H - 0.0816 O - 0.216 Si - 0.379 (mass ratio).

Measuring method:

Preparation issues from standard ER solutions whose activities were determined by suitable absolute method. Final control is based on gamma spectrometry on HPGe detector.

Note:

As the criterion of homogeneity standard deviation of the activity value of 1 cm³ element was chosen (n=10). The volume is calculated from the mass and the density. Pb-210 is in the radioactive equilibrium with Po-210.

Date of the certificate issue: 10.4.2015

Certificate validity 3 years

Customer:

Office fédéral de la santé publique OFSP

3003 Bern
SWITZERLAND




Control: Ing. Vlasta Zdybchová, RNDr. Pavel Dyrák, CSC.


Ing. Jiří Šuráň, MBA
Deputy director of JRB Prague


METROLOGY

Traceability refers to an unbroken chain of comparisons relating an instrument's measurements to a known standard.



Calibration to a traceable standard can be used to determine an instrument's bias, or accuracy.





METAS



NMI




USER


In many countries, national standards are maintained by a National Metrology Institute (NMI) → highest level of standards for the calibration / measurement traceability infrastructure in that country.

METROLOGY + RADIO = IONISING RADIATION



Traceability refers to an unbroken chain of comparisons relating an instrument's measurements to a known standard.



Calibration to a traceable standard can be used to determine an instrument's bias, or accuracy.



Radiometrology



Bq Gy Sv





Units of measure for radioactivity

The number of falling apples can be compared to the **Becquerel** (number of disintegrations per second).



The number of apples that hit the sleeper can be compared to the **Gray** (absorbed dose).

The effect on the body, based on the size or weight of the apples, can be compared to the **Sievert** (effective dose).


Source: C.E.A.



Radiometry



Bq **Gy** **Sv**





Units of measure for radioactivity

The number of falling apples can be compared to the **Becquerel** (number of disintegrations per second).


The number of apples that hit the sleeper can be compared to the **Gray** (absorbed dose).

The effect on the body, based on the size or weight of the apples, can be compared to the **Sievert** (effective dose).

Source: C.E.A.




Measurement Traceability




In many countries, national standards for IR are maintained by a designated institute.

In CH, DI = IRA

Gamma spectrometry standards




Gammaspectrometry standards




Mother/master solution

Standards



certificate



Gravimetric dilutions

Gravimetric procedures for radionuclide metrology are described in a monographie of the BIPM (1975)

- It is maybe old, but nothing change in chemical interactions.
- Detailed description of precautions (chemistry, physics, ...)
- One important recommendation is that the dilution factor must be < 50 and preferably > 30 .

PROCEDURES FOR ACCURATELY DILUTING
AND DISPENSING RADIOACTIVE SOLUTIONS



1975

Example: procedure to produce Co-57 standard

Goal of the manipulation:

Realize three vials of Co-57 with an activity of around 200 MBq (concentration 40 MBq/g) from a mother sample at high concentration (4 GBq/g)

Use of several samples to obtain an accurate measurement of the activity:

Traceable measurement of the activity concentration using a sealed ampoule (using the CIR)

Control the presence of impurities (Co-58, Co-56...) by gamma spectrometry

Exploitation de la source de Co-57 Eust-2012 pour la fabrication de sources de référence

Reference date : 12 mars 2012

Source preparation date : 13 mars 2012


Cobalt-57

Half-life : 6523.2 (12) h

Carrier : 25µg/g CoCl2 in HCl 0.1 M

Density CoCl2 : 0,999 g/cm3

Source Eust-2012
Concentration : 4.205 GBq/g
Mass ~ 0,2416 g
Activity : 1.016 GBq



Exploitation de la source de Co-57 Eust-2012 pour la fabrication de sources de référence

Reference date : 12 mars 2012


Source preparation date : 13 mars 2012

Cobalt 57

Half-life : 6523.2 (12) h

Carrier : 25µg/g CoCl₂ in HCl 0.1 M

Density CoCl₂ : 0,999 g/cm³



Source Eust-2012
Concentration : 4.205 GBq/g
Mass ~ 0.2416 g
Activity : 1.016 GBq

Mother solution in 2ml ampoule




*Tout prélever
Diluer à 24,5 g
(deux rinçages du
flacon original
inclus)*

1


F_{dil} ~ 101

M57Co28
Concentration : 41.469 MBq/g
Mass : 24.5 g
Activity : 1.016 GBq



Erlenmeyer 100 ml

6 Solde



M57Co28A3
C : 41.469 MBq/g
Mass : ~ 3 g
A : 124.408 MBq


2

4

*Prélever 500
ul (0.5 g)
diluer à 20ml
(20g)*



M57Co28F1, F2 & F3 vials
USP vials filled to 5 g
Concentration : 41.469 MBq/g
Mass : 5 g
Activity : 207.347 MBq



M57Co28A1 A2 ampoules
IER 5 filled to 3ml
Concentration : 41.469 MBq/g
Mass : 3 g
Activity : 124.408 MBq

F_{dil} ~ 40




M57Co29
FSPE Metas
C : 1.0368 MBq/g
M : 20 g
A : 20.735 MBq

F_{dil} ~ 400

*Prélever 50 ul
(50 mg)
Diluer à 20ml
(20g)*

5



M57Co30 F1 & F2
FSPE TB
C : 2.591 kBq/g
Mass : 20 g
A : ~ 51.8 kBq

Exploitation de la source de Co-57 Eust-2012 pour la fabrication de sources de référence

Reference date : 12 mars 2012

Source preparation date : 13 mars 2012

Cobalt-57

Half-life : 6523.2 (12) h

Carrier : 25µg/g CoCl₂ in HCl 0.1 M

Density CoCl₂ : 0,999 g/cm³


Source Eust-2012

Concentration : 4.205 GBq/g

Mass ~ 0.2416 g

Activity : 1.016 GBq

Multiple dilutions to reach a factor 100 using carrier CoCl₂ dans HCl 0.1 M



Exploitation de la source de Co-57 Eust-2012 pour la fabrication de sources de référence

Reference date : 12 mars 2012

Source preparation date : 13 mars 2012

Cobalt-57

Half-life : 6523.2 (12) h

Carrier : 25µg/g CoCl2 in HCl 0.1 M

Density CoCl2 : 0,999 g/cm3

Source Eust-2012
Concentration : 4.205 GBq/g
Mass ~ 0,2416 g
Activity : **1.016 GBq**

*Tout prélever
Diluer à 24,5 g
(deux rinçages du
flacon original
inclus)*

F_{dil} ~ 101

M57Co28
Concentration : 41.469 MBq/g
Mass : 24,5 g
Activity : **1.016 GBq**

Erlenmeyer 100 ml

6 Solde

M57Co28A3
C : 41.469 MBq/g
Mass : ~ 3 g
A : **124.408 MBq**

Taking 3 x 5 g with syringe to produce 3 vials (endproduct)

*Prélever 500
ul (0,5 g)
diluer à 20ml
(20g)*

F_{dil} ~ 40

M57Co29
FSPE Metas
C : 1.0368 MBq/g
M : 20 g
A : **20.735 MBq**


F_{dil} ~ 400

*Prélever 50 ul
(50 mg)
Diluer à 20ml
(20g)*

M57Co30 F1 & F2
FSPE TB
C : **2.591 kBq/g**
Mass : 20 g
A : ~ **51.8 kBq**

M57Co28F1, F2 & F3 vials
USP vials filled to 5 g
Concentration : 41.469 MBq/g
Mass : 5 g
Activity : **207.347 MBq**

M57Co28A1 A2 ampules
IER 5 filled to 3ml
Concentration : 41.469 MBq/g
Mass : 3 g
Activity : **124.408 MBq**



Exploitation de la source de Co-57 Eust-2012 pour la fabrication de sources de référence

Reference date : 12 mars 2012

Source preparation date : 13 mars 2012

Cobalt-57

Half-life : 6523.2 (12) h

Carrier : 25µg/g CoCl₂ in HCl 0.1 M

Density CoCl₂ : 0,999 g/cm³

Source Eust-2012

Concentration : 4.205 GBq/g

Mass ~ 0,2416 g

Activity : 1.016 GBq

*Tout prélever
Diluer à 24,5 g
(deux rinçages du
flacon original
inclus)*

F_{dil} ~ 101

M57Co28
Concentration : 41.469 MBq/g
Mass : 24,5 g
Activity : 1.016 GBq

Erlenmeyer 100 ml

6 Solde

M57Co28A3
C : 41.469 MBq/g
Mass : ~3 g
A : 124.408 MBq

*Prélever 500
ul (0,5 g)
diluer à 20ml
(20g)*

F_{dil} ~ 40

M57Co29
FSPE Metas
C : 1.0368 MBq/g
M : 20 g
A : 20.735 MBq

F_{dil} ~ 400


*Prélever 50 ul
(50 mg)
Diluer à 20ml
(20g)*

M57Co30 F1 & F2
FSPE TB
C : 2.591 kBq/g
Mass : 20 g
A : ~ 51.8 kBq

M57Co28A1 A2 ampoules
IER 5 filled to 3ml
Concentration : 41.469 MBq/g
Mass : 3 g
Activity : 124.408 MBq

M57Co28F1, F2 & F3 vials
USP vials filled to 5 g
Concentration : 41.469 MBq/g
Mass : 5 g
Activity : 207.347 MBq

Taking 2 x 3 g with syringe to produce 2 ampoules for CIR measurement
→ check of dilution factor



Exploitation de la source de Co-57 Eust-2012 pour la fabrication de sources de référence

Reference date : 12 mars 2012

Source preparation date : 13 mars 2012

Cobalt-57

Half-life : 6523.2 (12) h

Carrier : 25µg/g CoCl₂ in HCl 0.1 M


Density CoCl₂ : 0,999 g/cm³

Source Eust-2012

Concentration : 4.205 GBq/g

Mass ~ 0,2416 g

Activity : 1.016 GBq



Taking 0.5 g with pipette and complete at 20 g with carrier to perform gamma spectrometry
→ Check for impurities

Glassware preparation

- *Treated with deconnex (detergent) during 24h*
- *Rinsed and bathed with distilled water during 24h*
- *Rinsed and bathed with carrier during 24h*
- *Rinse and streamed during 8h*
- *Each container is referred and covered with parafilm*

Preparation of accessories

- *2 syringes 1ml, 1 syringe 5ml, 1 syringe 10 ml; 100 ml of carrier; 2 pipettes 500ml et 50ml + laboratory material, shielding and contamination monitor.*
- *Each syringe is refered*

Weighing room is climatized 24 h before use

At each dilution step the containers are weighed to know the mass content and to estimate the waste (leftovers in containers, syringes ...)

The weighing values are automatically saved in an Excel file and recorded by hands

The continuous mass recording is essential to know precisely the mass of the solution of each container

The final uncertainty on mass is < 0.01%

But how do we make/get the mother/master solution?




→ Magic of the realization of the unit Bq

Realization



Radioactive solution



Primary measurement

Primary measurement

Primary measurement

Defined solid angle alpha counting (Rn-222)



LSC (F-18, Y-90, beta), Triple to Double Coincidence Ratio (TDCR) (H-3, C-14, Ni-63, Fe-55...)

$4\pi\gamma$ -counting (F-18, I-123, Tc-99m)

Coincidence counting:


$4\pi\beta$ - $4\pi\gamma$ -counting (F-18, Cs-134, Ba-133,...)

$4\pi\beta\gamma$ -counting (Co-57, I-123, I-125, I-131, Tc-99m)



β - γ coincidence counting


For a radionuclide which emits a beta (β) followed by a gamma (γ) particle, one can write :



$$R_\beta = R_0 \times \epsilon_\beta$$

$$R_\gamma = R_0 \times \epsilon_\gamma$$

$$R_c = R_0 \times \epsilon_\beta \times \epsilon_\gamma$$



$$R_0 = \frac{R_\beta R_\gamma}{R_c}$$


Realization



Conservation



Radioactive
solution




$C_{A, \text{mes}}$


$C_{A, \text{mes}} \leftrightarrow A_e$



Realization




Radioactive
solution



Conservation



$$C_{A, \text{mes}} \leftrightarrow A_e$$



$$C_{A, \text{mes}}$$

Political Map of the World, April 2006






Primary laboratory missions

- Realization of the physical quantity
- Conservation
- Dissemination

Realization



Conservation




Up to now, **30** international comparisons of radionuclides, of which 17 are now CCRI(II) key comparisons → **this is it for now!**

Traceability brought to clinical site → Example of radionuclide calibrators


IRA



Primary standard
IRA (s=0.3%)



Reference
radioactive
solutions





CIR, S= 0.31%





METAS
Verification
 $\Delta < 10\%$ (γ)



Building a chain of traceability





Building a chain of traceability




Must take into account:
goal,
technical limitations,
financial limitations,
....

Building a chain of traceability



Building a chain of traceability



Merci de votre attention

Danke für Ihre Aufmerksamkeit

Thanks you.....

